

DECLARATION OF CONFORMITY

(in accordance with ISO/IEC 17050-1)

We declare, under our sole responsibility, that the product listed below conforms to the provisions of:

- the European Council Directive 93/42/EEC of 14 June 1993 concerning medical devices
- the Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment (RoHS)

Manufacturer's Name and Business Address:	Welch Allyn, Inc. 4341 State Street Road Skaneateles Falls, NY 13153 USA
EC REP	Regulatory Affairs Representative Welch Allyn Limited Navan Business Park, Dublin Road Navan, Co. Meath C15 AW22 Ireland
Product Name ^{1,2} :	HScribe 5 HScribe 6 HScribe 6 Workstation HScribe 6 Server Vision Express
REF _{1,2}	The HScribe and Vision Express comes in different configurations HSCRIBE-5XX-XXXXX Where "X" can be a letter from A to Z representing the following: HSCRIBE- [5] [Model] [Configuration] - [Option1] [Option2] [Option3] [Option4] [Option5] HSCRIBE-6XX-XXXXX Where "X" can be a letter from A to Z representing the following: HSCRIBE- [6] [Model] [Configuration] - [Option1] [Option2] [Option3] [Option4] [Option5] HSCRIBWS-6XX-XXXXX Where "X" can be a letter from A to Z representing the following: HSCRIBWS - [6] [Model] [Configuration] - [Option1] [Option2] [Option3] [Option4] [Option5] HSERV-6XX-XXXXX Where "X" can be a letter from A to Z representing the following: HSERV - [6] [Model] [Configuration] - [Option1] [Option2] [Option3] [Option4] [Option5] BURV53H-X Where "X" can be a letter from A to Z representing the following: BURV53H - [Option1]

¹ applicable to the medical device directive, 93/42/EEC

² applicable to the RoHS directive, 2011/65/EU

<div style="border: 1px solid black; display: inline-block; padding: 2px;">#</div> ^{1,2}	901143 – HOLTER ANALYSIS SOFTWARE	
Medical Device Conformity Assessment Route Annex ¹ :	II	
Medical Device Classification ¹ :	IIa	
Medical Device Classification Rules ¹ :	10	
GMDN Code and Term ¹ :	41651 - Electrocardiographic long-term ambulatory recording analyser application software	
Notified Body ¹ : (CE 0459)	LNE G-MED, 1, rue Gaston Boissier 75015 Paris France EC-certificate No. 35913	
Standards Applied	Number	Title
	EN/IEC 62304 ¹	Medical Device Software – Software Life Cycle Processes
	EN/IEC 60601-1 ¹	Medical Electrical Equipment – Part 1: General Requirements for Basic Safety and Essential Performance
	EN/IEC 60601-1-2 ¹	Medical electrical equipment - Part 1-2: General Requirements for Safety – Collateral Standard: Electromagnetic Compatibility – Requirements and Tests
	EN/IEC 62366 ¹	Medical devices – Application of Usability Engineering to Medical Devices
	EN/IEC 60601-1-6 ¹	Medical Electrical Equipment – Part 1-6: General Requirements for Safety – Collateral Standard: Usability
	EN/IEC 60601-2-47 ¹	Medical electrical equipment - Part 2-47: Particular requirements for the safety, including essential performance, of ambulatory electrocardiographic systems

¹ applicable to the medical device directive, 93/42/EEC

² applicable to the RoHS directive, 2011/65/EU

Authorised Signatory:



07-05-2019

Mark Elliott,
Director Quality Assurance

Date

Milwaukee, Wisconsin, USA
Place of Issue

¹ applicable to the medical device directive, 93/42/EEC

² applicable to the RoHS directive, 2011/65/EU